



CheKine™ Micro Glutathione Peroxidase (GSH-Px) Activity Assay Kit

Cat #: KTB1640

Size: 48 T/96 T

	Micro Glutathione Peroxidase (GSH-Px) Activity Assay Kit		
REF	Cat #: KTB1640	LOT	Lot #: Refer to product label
	Applicable samples: Serum, Plasma, Animal Tissues, Cells, Bacteria		
	Storage: Stored at -20°C for 12 months, protected from light		

Assay Principle

Glutathione Peroxidase (GSH-Px) is one of the main enzymes that catalyze the oxidation of reduced glutathione (GSH) in the glutathione redox cycle. GSH-Px not only specifically catalyzes the reaction between reduced glutathione and ROS to form oxidized glutathione GSSG, thereby protecting the biofilm from ROS damage and maintaining the normal function of cells; it also protects the liver and improves the body immunity, antagonizing the harmful effects of harmful metal ions on the body and increasing the body's ability to resist radiation. CheKine™ Micro Glutathione Peroxidase (GSH-Px) Activity Assay Kit provides a simple method for detecting activity of GSH-Px in a variety of biological samples such as Serum, Plasma, Animal Tissues, Cells and Bacteria. GSH-Px catalyzes H₂O₂ to oxidize GSH to produce GSSG; Glutathione Reductase (GR) catalyzes NADPH to reduce GSSG to regenerate GSH, while NADPH oxidizes to produce NADP⁺; NADPH has a characteristic absorption peak at 340 nm, while NADP⁺ does not; NADPH dehydrogenation rate is determined by measuring the decrease rate of absorbance at 340 nm to calculate GSH-Px activity.

Materials Supplied and Storage Conditions

Kit components	Size		Storage conditions
	48 T	96 T	
Assay Buffer	50 mL+20 mL	100 mL+20 mL	4°C
Substrate	1	1	-20°C, protected from light
GR	10 µL	10 µL	4°C, protected from light
H ₂ O ₂	200 µL	200 µL	4°C, protected from light

Materials Required but Not Supplied

- Microplate reader or ultraviolet spectrophotometer capable of measuring absorbance at 340 nm
- Incubator
- 96-well UV plate or microquartz cuvette, precision pipettes, disposable pipette tips

- Freezing centrifuge
- Deionized water
- Dounce homogenizer (for tissue samples)

Reagent Preparation

Assay Buffer: Ready to use as supplied. Equilibrate to room temperature before use. Store at 4°C.

Working Substrate: Prepare before use, add 20 mL Assay Buffer to dissolve Substrate, mix well; Store and aliquot the remaining reagent at -20°C for 1 month, protected from light.

GR: Ready to use as supplied. Equilibrate to room temperature before use; Store at 4°C, protected from light.

Working H₂O₂: Prepare before use, add 21.5 μL H₂O₂ and 5 mL deionized water, mix well and use the prepared reagent on the same day; Store at 4°C, protected from light.

Working Reagent: Prepare before use, According to the number of samples, the ratio of adding 1 μL GR to 2 mL Working Substrate is prepared, mix well (run out the same day); Working Reagent is preheated to 25°C (general species) or 37°C (mammal) for 30 min.

Sample Preparation

Note: Fresh samples are recommended, if not assayed immediately, samples can be stored at -80°C for one month.

1. Tissue Samples: Wash Tissue with cold PBS to remove blood as much as possible. Weigh 0.1 g Tissues, add 1 mL Assay Buffer and homogenize on ice. Centrifuge at 10,000 g for 10 min at 4°C. Use supernatant for assay, and place it on ice to be tested.

2. Serum and other liquids: direct determination. If necessary, can be diluted with physiological saline and the result should be multiplied by the dilution factor.

3. Bacteria, Cells: Collect 5×10⁶ cells or bacteria into the centrifuge tube, wash Cells or Bacteria with cold PBS, discard the supernatant after centrifugation, add 1 mL Assay Buffer to ultrasonically disrupt the cells or bacteria 5 min (power 20% or 200 W, ultrasonic 3 s, interval 7 s, repeat 30 times). Centrifuge at 10,000 g for 10 min at 4°C. Use supernatant for assay, and place it on ice to be tested.

Note: The whole processes need to be carried out on ice, and the enzyme activity should be determined on the same day, to avoid repeated freeze-thaw of the homogenate solution. In the detection of GSH-Px activity in Cells, the Cell number must be between 3-5×10⁶, and the extraction of GSH-Px in cells can be followed by Assay Buffer grinding or ultrasonic treatment, Cells cannot be treated by lysate. For additional measurement, it is recommended to use Abbkine Protein Quantification Kit (BCA Assay) (Cat #: KTD3001).

Assay Procedure

1. Preheat the microplate reader or ultraviolet spectrophotometer for more than 30 min, and adjust the wavelength to 340 nm. Visible spectrophotometer was returned to zero with deionized water.

2. Working Regent place at 25°C (for general species) or 37°C (for mammals) incubation for 30 min.

3. Add the following reagents to the 96-well UV plate or microquartz cuvette, follow the table below to set the test:

Reagent	Blank Well (μL)	Test Well (μL)
Deionized Water	20	0
Sample	0	20
Working Reagent	160	160
Working H ₂ O ₂	20	20

4. Mix well. The absorbance values were measured at 340 nm for 10 s and 70 s. The absorbance value of blank well 10 s is A₁, 70

s is A₂; the absorbance value of test well 10 s is A₃, 70 s is A₄, calculate $\Delta A_{\text{Blank}}=A_1-A_2$, $\Delta A_{\text{Test}}=A_3-A_4$.

Note: Blank Well only need to measure 1-2 times. In order to guarantee the accuracy of experimental results, need to do a pre-experiment with 2-3 samples. The reaction temperature has influence on the result. Keep the temperature at 25°C (for general species) or 37°C (for mammals).

Data Analysis

Note: We provide you with calculation formulae, including the derivation process and final formula. The two are exactly equal. It is suggested that the concise calculation formula in bold is final formula.

A. Use 96-well UV plate calculation formula as below:

(1) By protein concentration

Active unit definition for GSH-Px: at 25°C or 37°C, 1 μmol NADPH oxidation per milligram of protein per minute was catalyzed.

$$\text{GSH-Px activity (U/mg prot)} = [(\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div (\epsilon \times d) \times V_{\text{Total}} \times 10^6] \div (\text{Cpr} \times V_{\text{Sample}}) \div T = \mathbf{3.216 \times (\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div \text{Cpr}}$$

(2) By sample fresh weight

Active unit definition for GSH-Px: at 25°C or 37°C, 1 μmol NADPH oxidation per gram of sample per minute was catalyzed.

$$\text{GSH-Px activity (U/g)} = [(\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div (\epsilon \times d) \times V_{\text{Total}} \times 10^6] \div (V_{\text{Sample}} \div V_{\text{Sample Total}} \times W) \div T = \mathbf{3.216 \times (\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div W}$$

(3) By number of cells or bacteria

Active unit definition for GSH-Px: at 25°C or 37°C, 1 μmol NADPH oxidation per 10⁴ cells or bacteria of sample per minute was catalyzed.

$$\text{GSH-Px activity (U/10}^4) = [(\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div (\epsilon \times d) \times V_{\text{Total}} \times 10^6] \div (500 \times V_{\text{Sample}} \div V_{\text{Sample Total}}) \div T = \mathbf{3.216 \times (\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div 500}$$

(4) By liquid volume

Active unit definition for GSH-Px: at 25°C or 37°C, 1 μmol NADPH oxidation per mL of liquid per minute was catalyzed.

$$\text{GSH-Px activity (U/mL)} = [(\Delta A_{\text{Test}} - \Delta A_{\text{Blank}}) \div (\epsilon \times d) \times V_{\text{Total}} \times 10^6] \div V_{\text{Sample}} \div T = \mathbf{3.216 \times (\Delta A_{\text{Test}} - \Delta A_{\text{Blank}})}$$

Where: $\Delta A_{\text{Blank}} = A_1 - A_2$, $\Delta A_{\text{Test}} = A_3 - A_4$; ϵ : NADPH molar extinction coefficient 6.22×10^3 L/mol/cm; d: 96-well UV plate light path, 0.5 cm; V_{Total} : Total volume of reaction system, 200 μL = 2×10^{-4} L; 10^6 : 1 mol = 1×10^6 μmol; Cpr: Protein concentration of supernatant (mg/mL); W: sample mass, g; V_{Sample} : Volume of supernatant added to the reaction system, 20 μL = 2×10^{-2} mL; $V_{\text{Sample Total}}$: Volume of extraction solution, 1 mL; T: Reaction time, 1 min; 500: Number of cells or bacteria, 5×10^6 .

B. Microquartz cuvette calculation formula

The optical diameter d:0.5 cm in the above calculation formula can be adjusted to d:1 cm for calculation.

Recommended Products

Catalog No.	Product Name
KTB1600	CheKine™ Micro Reduced Glutathione (GSH) Assay Kit
KTB1610	CheKine™ Micro Glutathione Oxidized (GSSG) Assay Kit
KTB1620	CheKine™ Micro Glutathione Reductases (GR) Activity Assay Kit
KTB1630	CheKine™ Micro Glutathione S-Transferase (GST) Assay Kit
KTB1650	CheKine™ Micro Thioredoxin Reductase (TrxR) Assay Kit
KTB1660	CheKine™ Micro Thioredoxin Peroxidase (TPX) Assay Kit

Disclaimer

The reagent is only used in the field of scientific research, not suitable for clinical diagnosis or other purposes.

